



Showerbuddy Global Ltd.

PO Box 494,  
Waikanae, 5250  
New Zealand

25<sup>th</sup> February 2013

Dear Barry,

Reference: testing the Showerbuddy Global Limited SB1Mobile shower chair, our Laboratory reference is 491832.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method: AS/NZS 3973:2009 *Shower / toilet chairs, Mobile and Static* (except materials and surface finishes).

The official nine (9) page report, dated 25<sup>th</sup> February 2013 and referencing job number 491832, identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standards are indicated in the report.

The Showerbuddy Global Limited SB1 ShowerBuddy Mobile shower chair, as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of AS/NZS 3973:2009 *Shower / toilet chairs, Mobile and Static* (excluding the methods indicated in the report as "not tested") for user mass 160 kg.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel  
Senior Test Technician, Test Laboratory  
Authorised signatory



**NovitaTech (a division of Novita Children's Services)**  
171 Days Road, Regency Park, South Australia 5010  
PO Box 2438, Regency Park, South Australia 5942  
Telephone 1300 855 585 • Fax (08) 8243 8208  
enquiries@novita.org.au • www.novitatech.org.au  
ABN 53 137 537 636

